

Enterococcus faecium

(Orla-Jensen 1919) Schleifer and Kilpper-Bälz 1984

CECT 410; WDCM 00010

La **Cepa Certificada CECT** es un material de referencia recomendado por Normas Internacionales de Calidad como microorganismo control y que ha sido objeto de ensayo en la CECT con los resultados que se muestran en este certificado.

CECT Certified Strain is a reference material recommended as testing microorganism in International Quality Standards, and assayed at CECT with the results shown in this certificate.

CHARACTERIZATION


Miniaturized biochemical test rapid ID 32 STREP

+	+	*	*	-	-	+	+	-	+	+	-	+	*	+	+	*	*
1	2	4	1	2	4	1	2	4	1	2	4	1	2	4	1	2	4
ADH	βGLU	βGAR	βGUR	αGAL	PAL	RIB	MAN	SOR	LAC	TRE	RAF	VP	APPA	βGAL	PYRA	βNAG	GTA
3 / 7			0 / 1			3			3			5 / 7			1 / 3 / 5		

*	-	-	+	-	-	+	+	-	*	-	+	+	-
1	2	4	1	2	4	1	2	4	1	2	4	1	2
HIP	GLYG	PUL	MAL	MEL	MLZ	SAC	LARA	DARL	MBDG	TAG	βMAN	CDEX	URE
0 / 1			1			3			4 / 5			1	

Identification: 89.5-99.9% ID *Enterococcus faecium* ^{*This result may vary}

Growth on selective/differential culture media



Bile Aesculin Agar
 Blackening of medium around colonies



Slanetz-Bartley Agar (with TTC solution 1%)
 Red or brown colonies



KF Agar (with TTC solution 1%)
 Red or pink colonies, surrounded by a yellow zone